


## CERTIFICATE OF ANALYSIS

Product Name	FOXO4-DRI peptide, Cat.#: 318716 (TFA removed)
Lot No.	MTFA31472-0613
Sequence	ltlrkpeaseiaqsileaysqngwanrrsggkrppprrrqrkkrg (Acetate Salt)
Dissolution condition	100%H <sub>2</sub> O
Length	46AA
Modification	allD-AA, Acetate Salt
Molecular Weight (MW)	5358.05
Storage	-20°C

Test Items	Specifications	Results
MW by MS	5358.85	Conforms
Purity by HPLC	>95%	96.535%
Peptide content	N/A	N/A
Moisture content	N/A	N/A
Acetic acid content	N/A	N/A
Appearance	White to off-white lyophilized powder	Conforms
Quantity	1g	5.0mg*140+10.0mg*30

NovoPro Bioscience Inc. (hereafter NovoPro) warrants material of said quality at the time of sale. It is the sole responsibility of the customers to determine the adequacy of all materials for any intended or specific purpose or use. NovoPro's sole obligation is to replace the material up to the extent of the purchase price. This warranty applies only to products in original packaging and does not apply to a product which has been tampered with or altered in any way in or which has been misused or damaged by accident or negligence. All claims must be received writing (by fax or email) within 30 days from date when product arrive at the destination city and failure to do so shall constitute a waiver by customers for any and all such claims.

Certified by:



Quality Assurance Department

Jun/19/2024






### Sample Information

Order ID :MTFA31472  
 Name :FOXO4-DRI  
 Sequence :ltlrkepaseiaqsileaysqngwanrrsggkrppprrrqrrkrg (Acetate Salt)  
 Lot. No :MTFA31472-0613  
 Pump A :0.1%Trifluoroacetic in 100% water  
 Pump B :0.1%Trifluoroacetic in 100% acetonitrile  
 Total Flow :1ml/min  
 Wavelength :214nm  
 Analytical column type :Kromasil 100-5-C18 (4.6\*250mm\*5 µm)  
 Dissolution method :100%H<sub>2</sub>O  
 Inj. Volume : 30ul  

Time	Module	Action	Value
0.01	Pumps	B.Conc	20
30.00	Pumps	B.Conc	50
33.00	Pumps	B.Conc	100
38.00	Pumps	B.Conc	100
40.00	Pumps	B.Conc	20
50.00	Controller	Stop	

### Chromatogram

mV




### Peak Table

Detector A 214nm

Peak#	Ret. Time	Area	Height	Area%
1	14.571	199232	7516	2.376
2	14.860	8093449	528857	96.535
3	15.188	91279	16306	1.089
Total		8383961	552679	100.000



# MS Spectrum



## Sample Information

Dissolution method	:15% ACN+85% H <sub>2</sub> O	Interface	:ESI	Prerod Bias	:+4.5kv
Date Acquired	:2024/06/17 13:16:39	Nebulizing Gas Flow	:1.50L/min	Detector	:-0.2kv
Injection Volume	:0.2ul	CDL Temp	:250°C	T.Flow	:0.2ml/min
Block Temp	:200	CDL Volt	:0v	B.conc	:50% H <sub>2</sub> O/50% MEOH
Order ID	:MTFA31472				
Name	:FOXO4-DRI				
Sequence	:ltlrkaseiaqsileaysqngwanrrsggkrppprrrrkrkrg (Acetate Salt)				
Lot.No	:MTFA31472-0613				
Theoretical	:5358.05				
Observed	:5358.85				